

5. Specificity

The specificity of the Cannabinoids ELISA for was determined by generating inhibition curves for each of the compounds listed below

Compound	prox. ng/ml equivalent to 10 ng THCA/ml	Cross-reactivities
11-nor-9-carboxy- Δ^8 -THC	11	110
Δ^9 -THC	48	21
Δ^8 -THC	22	45
11-hydroxy- Δ^9 -THC	>1000	<5
8-11-Dihydroxy Δ^9 -THC	>1000	<5
annabitol	>1000	<5
Cannabidiol	>1000	<5

6. Cross-Reactivity with Unrelated Drugs

Aliquots of a human urine matrix were spiked with the following compounds at a concentration of 2,000 ng/ml. None of these compounds gave values in the assay that were equal to or greater than the assay sensitivity level.

Acetaminophen, Acetylsalicylic acid, Amphetamine, Aminopyrine, Ampicillin, Amobarbital, Ascorbic acid, Atropine, Barbitol, Butabarbital, Caffeine, Cocaine, Carbamazepine, Codeine, Chloroquine, Chlorpromazine, Carbronal, Desipramine, Dextromethorphan, Dextropropoxyphene, 5,5-Diphenylhydantoin, 10-11-Dihydrocarbamazepine, Diazepam, Ethosuximide, Estriol, Estrone, Estradiol, Ethotoin, Glutethimide, Hexobarbital, Ibuprofen, Imipramine, Lidocaine, LSD, Methadone, Methadone-primary metabolite, Methaqualone, Methamphetamine, Metharbital, Mephentoin, β -Methyl- β -propylsuccinimide, Mephobarbital, Methyl PEMA, Methsuximide, 4-Methylprimidone, Morphine, Meperidine, Niacinamide, Norethindrone, N-Normethsuximide, Phenobarbital, Phensuximide, PEMA, Primidone, Phencyclidine, Pentobarbital, Phenothiazine, Phenylpropanolamine, Procaine, Quinine, Secobarbital, Tetracycline, Tetrahydrozoline

REFERENCES

1. Urine Testing for Drugs of Abuse, National Institute on Drug Abuse Research Monograph . 73, 1986.
2. Rodgers, R., Crowl C.P., Eimstad W.M., et al.: Homogeneous enzyme immunoassay for cannabinoids in urine. Clin. Chem. 24: 95 (1978).
3. Teale, J.D., Foxman, E.J., King, L.S., Pial, E.M. and Marks, V.: The development of a radioimmunoassay for cannabinoids in blood and urine. J. Pharm. Pharmacol. 27: 465 (1975).
4. Mule, S.J., Lomax, P. and Gross, S.J.: Active and passive marijuana exposure tested by three immunoassays and GC/MS in urine. J. Anal. Toxicol. 12: 113 (1988).
5. Cone, E.J. and Johnson, R.E.: Contact highs and urinary cannabinoid excretion after passive exposure to marijuana smoke. Clin. Pharmacol. Ther. 40: 247 (1986).

2010-01-25 (Mfg: 052001)

Warning

All of Bio-Quant ELISA Kits have not been tested for clinical use and are not approved in the United States by the FDA for diagnostic clinical use. They are components or reagents made solely for research use, further manufacturing and export use. It is the commitment of Bio-Quant customers to receive its products solely for the purpose of exportation or research, and not for the purposes of clinical diagnostic use.

BIO-QUANT DOES NOT MAKE ANY OTHER WARRANTY OR REPRESENTATION WHATSOEVER, WHETHER EXPRESS OR IMPLIED, WITH RESPECT TO THESE PRODUCTS. IN PARTICULAR Bio-Quant, INC. DOES NOT MAKE ANY WARRANTY OF SUITABILITY, NON-INFRINGEMENT, MERCHANTABILITY OR FITNESS FOR ANY PARTICULAR PURPOSE OF ANY PRODUCT



Cannabinoids (THCA/CTHC) ELISA Kit

Catalog No. BQ 205-096 (96 TESTS)

INTENDED USE

The Cannabinoids Direct ELISA Kit provides only a preliminary analytical test result. A more specific alternate chemical method must be used in order to obtain a confirmed analytical result. Gas chromatography/ mass spectrometry (GS-MS) is the preferred confirmatory method (1). Professional judgment should be applied to any drug of abuse test result, particularly when preliminary positive results are used.

EXPLANATION OF THE TEST

The Cannabinoids Direct ELISA Kit is a specific and sensitive in-vitro test to detect the presence of cannabinoids in samples such as whole blood, serum, plasma and urine. Δ^9 -THC (a member of the cannabinoid family) is the primary psychoactive ingredient of marijuana (1). Cannabinoid metabolites appear in urine two to four hours after a marijuana smoke and may persist for days (up to thirty) (1- 3). Thus a urine assay reasonably serves to detect cannabis use even though a considerable period may have elapsed since smoking or ingestion of marijuana.

PRINCIPLES OF THE PROCEDURE

The Cannabinoids Direct ELISA Kit is based upon the competitive binding to antibody of enzyme labeled antigen and unlabeled antigen, in proportion to their concentration in the reaction mixture. A 10 μ l. aliquot of a diluted unknown specimen is incubated with a 100 μ l. dilution of enzyme (Horseradish peroxidase) labeled carboxy THC (THCA) derivative in micro-plate wells, coated with fixed amounts of oriented high affinity purified polyclonal antibody. The wells are washed thoroughly and a chromogenic substrate added. The color produced is stopped using a dilute acid stop solution and the wells read at 450 nm. The intensity of the color developed is inversely proportional to the concentration of drug in the sample. The technique is sensitive to 1 ng/ml. The THC Direct ELISA Kit avoids extraction of urine or blood sample for measurement. It employs a polyclonal high affinity, purified carboxy THC antibody. Due to the proprietary method of orienting the antibody on the polystyrene micro-plate much higher sensitivity is achieved compared to passive adsorption. This results in extremely small sample size reducing matrix effects and interference with binding proteins(s) or other macromolecules.

MATERIALS PROVIDED	96 tests
1. Microwell coated with polyclonal anti-carboxy THC	12x8x1
2. THC-Conjugate	12.5 ml
3. Immunalysis Positive Reference Standard	1 ml
4. Negative Standard	1 ml
5. TMB Substrate	14 ml
6. Stop Reagent	12.5 ml

MATERIALS NOT PROVIDED

1. Distilled or deionized water
2. Precision pipettes
3. Disposable pipette tips
4. ELISA reader capable of reading absorbance at 450nm
5. Absorbance paper or paper towel
6. Graph paper

STORAGE AND STABILITY

1. Store the kit at 2 - 4° C.
2. Keep microwells sealed in a dry bag with desiccants.
3. The reagents are stable until expiration of the kit.
4. Do not expose reagents to heat, sun, or strong light.

WARNINGS AND PRECAUTIONS

1. Potential biohazardous materials:
The calibrator and controls contain human source components, which have been tested and found non-reactive for hepatitis B surface antigen as well as HIV antibody with FDA licensed reagents. However, as there is no test method that can offer complete assurance that HIV, Hepatitis B virus or other infectious agents are absent, these reagents should be handled at the Biosafety Level 2, as recommended in the Centers for Disease Control/National Institutes of Health manual, "Biosafety in Microbiological and Biomedical Laboratories." 1984
2. This test kit is designed for Research Use Only.
3. Do not add sodium azide to samples as preservative. Do not use external controls containing sodium azide.
4. Viscous samples should always be diluted in phosphate buffered saline or distilled water prior to pipetting Do not pipette by mouth. Do not smoke, eat, or drink in the areas in which specimens or kit reagents are handled.
5. The components in this kit are intended for use as an integral unit. The components of different lots should not be mixed.
6. It is recommended that serum samples be run in duplicate.
7. Optimal results will be obtained by strict adherence to this protocol. Accurate and precise pipetting, as well as following the exact time and temperature requirements prescribed are essential. Any deviation from this may yield invalid data.

SPECIMEN COLLECTION**1. Precautions**

The Cannabinoids Direct ELISA Kit is to be used with human samples such as whole blood, serum, urine and plasma. has not tested all possible applications of this assay. The Cutoff criteria are important in deciding the sample dilution. It is recommended to dilute most blood samples either 1:5 or 1:10 depending on the cutoff used by the laboratory.

2. Additives

Specimens to which sodium azide has been added affect the assay.

3. Storage and Handling Instructions

Urine samples should be stored at 2-4°C until use. Samples should be well mixed before assay. Repeated freezing and thawing should be avoided. Urine samples should be shipped refrigerated with Blue Ice or equivalent.

ASSAY PROCEDURE

Bring all specimens and kit reagents to room temperature (18-26 °C) and gently mix.

1. Dilute specimens, to the necessary range with Phosphate Buffer Saline pH 7.0. (Urine samples are normally diluted 1:10 for a THCA cutoff of 50 ng/ml.) The dilution factor and volume added can be adjusted based on the laboratory's cutoff.
2. Add 10 µl. of appropriately diluted calibrators and standards to each well in duplicate.
3. Add 10 µl. of the diluted specimens in duplicate (recommended) to each well.
4. Add 100 µl. of the Enzyme Conjugate to each well. Tap the sides of the plate holder to ensure proper mixing.
5. Incubate for 60 minutes at room temperature (18-26° C) preferably in the dark, after addition of enzyme conjugate to the last well.
6. Wash the wells 6 times with 350 µl. distilled water using either a suitable plate washer or wash bottle taking care not to cross contaminate wells. If testing samples, containing abnormally high amounts of hemoglobin (some Postmortem samples), use 10 mM Phosphate buffered saline pH 7.0-7.4. This will lower potential nonspecific binding of hemoglobin to the well, thus lowering background color.
7. Invert wells and vigorously slap dry on absorbent paper to ensure all residual moisture is removed. This step is critical to ensure that residual enzyme conjugate, does not skew results. If using an automated system, ensure that the final aspiration on the wash cycle aspirates from either side of the well.
8. Add 100 µl. of Substrate reagent to each well and tap sides of plate holder to ensure proper mixing.

9. Incubate for 30 minutes at room temperature, preferably in the dark.
10. Add 100 µl. of Stop Solution to each well, to change the blue color to yellow.
11. Measure the absorbance at a dual wavelength of 450 nm and 650 nm.
12. Wells should be read within 1 hour of yellow color development.

The following data represent a typical dose/response curve.

CTHC ng/ml	Absorbance
0	1.985
2	1.413
5	0.955
10	0.751

The dose/response curve shown above should not be used in assay calculations. It is recommended that at least one in-house positive quality control sample be included with every assay run.

A dose response curve or a cutoff calibrator should be run with every plate.

RESULTS

If the average sample absorbance is equal to or less than the average absorbance of the laboratory THCA/CTHC positive reference standard the sample is **POSITIVE** for cannabinoids. If the average sample absorbance is greater than the average absorbance of the laboratory THCA/CTHC positive reference standard the sample is called **NEGATIVE** for cannabinoids.

Alternatively a dose response curve can be established by plotting standard concentration (abscissa) against corresponding absorbance (ordinate). Values for unknown samples are obtained by interpolation from the curve.

SPECIFIC PERFORMANCE CHARACTERISTICS**1. Accuracy**

35 whole blood samples and 60 urine samples collected from presumed non-users were tested in the Cannabinoids Direct ELISA Kit. One hundred percent of these normal samples measured negative at 20 ng/ml of THCA equivalents for whole blood and 50 ng/ml of THCA equivalents for urine. Forty whole blood samples which were previously confirmed positive for cannabinoids by GC-MS employing a cut-off of 10 ng/ml THCA were tested in the Cannabinoids Direct ELISA Kit. All the samples were found to be positive i.e. above the cut-off of 20 ng/ml.

2. Precision

The precision of the Cannabinoids Direct ELISA Kit has been verified by assessment of the mean, standard deviation (SD) and coefficients of variation (CV) in data resulting from repetitive assays.

3. Intra-assay Precision

Intra-assay precision was determined with reference controls.

A 0,2,5 and 10 ng/ml standard was assayed eight times in the same assay.

carboxy THC (ng/ml)	Mean Abs.	S.D.	C.V.%
0	1.905	0.139	7.3
2	1.114	0.103	9.4
5	0.752	0.066	8.8
10	0.549	0.042	7.7

4. Sensitivity

Assay sensitivity based on the minimum THCA concentration required to produce a four standard deviation from assay zero dose response (A₀) is 1 ng/ml.